



## **Instructions for Use Reprocessed (Reusable) Catheter Interface Cables**

### ***Reprocessed Device for Single Use***

**Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.**

#### **DEVICE DESCRIPTION**

The catheter interface cables are designed as electrode cables with a multi-pin connector on the distal end which connects to a catheter (diagnostic or ablation) and a multi-pin connector on the proximal end which connects to the appropriate equipment. The cables either interface a catheter with the appropriate external stimulation or recording equipment or, serve as an extension cable between a catheter and equipment out of immediate reach.

For a list of compatible generators, please refer to the applicable Original Manufacturer's Compatibility Matrix.

#### **INDICATIONS FOR USE**

The cable provides a means to interface an electrophysiology catheter to the appropriate equipment.

The cable may be re-used subject to the cleaning and sterilization services provided by Innovative Health.

#### **CONTRAINDICATIONS FOR USE**

- There are no known contraindications to this cable.

#### **WARNINGS**

- Patient or operator injury can result from improper handling or storage of the cable.
- If a break occurs in the cable wire or the cable becomes otherwise electrically discontinuous, arcing may occur in the patient-return or active circuit and may burn the patient or create a fire.
- Failure to isolate unused connector pins may result in accidental current pathways to the heart.
- Misconnection of the pins to the sockets could lead to malfunction of the catheter.

#### **PRECAUTIONS**

- Do not immerse the cable connectors in fluids.
- If used in the presence of electrical equipment, noise may be induced into signals conducted by the cable.
- For reuse of the cable, return to Innovative Health for additional processing.
- Do not expose cable to organic solvents.
- Inspect the packaging and cable for damage or defects (i.e. cuts, kinks, nicks, crushed or elongated sections) prior to use.
- Use patient isolated equipment.

#### **ADVERSE REACTIONS**

None.

## ***Instructions for Use: Reprocessed (Reusable) Catheter Interface Cables***

### **DIRECTIONS**

- The package label is detachable and may be affixed to the medical record of the patient.
- Before beginning the procedure, verify compatibility of all devices and accessories.
- **Note:** When connecting to a radio-frequency generator, refer to the appropriate Original Manufacturer's generator instructions for details and system compatibility.
- Inspect the cable and packaging before opening. The contents of the package are sterile unless the package is opened or damaged. If the cable is damaged or the packaging is compromised, do not use the cable. Do not attempt to repair any damage. Return the cable and packaging to Innovative Health.
- Do not attempt to resterilize.
- Ensure alignment of pins via rotation as required, prior to connection.
- Ensure each cable end is securely connected to the proper equipment. The "system" end of the cable states "To System" and must be connected to the System. The "catheter" end of the cable states the catalog number and must be connected to the catheter.
- For Boston Scientific SureLink cables, an audible click indicates that the two connectors are locked together.
- For information on how to connect a specific cable, refer to the applicable catheter and equipment instructions for use.

**Instructions for Use: Reprocessed (Reusable) Catheter Interface Cables**

**EXPLANATION OF SYMBOLS**



Federal Law in the USA restricts this device to sale by or on the order of a physician



Sterilized by Ethylene Oxide Gas



Catalog Number



Serial Number



Lot Number



Use by Date



Do Not Reuse



Do Not Resterilize



Consult Instructions for Use



Do Not Use if Package is Damaged



Keep Product Dry



Keep Away from Sunlight

As the reprocessor, Innovative Health is solely responsible for this device. All Original Manufacturer (OM) information is provided for device identification and may contain trademarks from third parties that do not sponsor this device.

**Sterilization:** This product and its packaging have been sterilized with ethylene oxide (EO) gas. Even though the product is processed in compliance with all applicable laws and regulations relating to EO exposure, Proposition 65, a State of California voter initiative, requires the following notice:

**Warning:** This product and its packaging have been sterilized with EO. The packaging may expose you to EO, a chemical known to the State of California to cause cancer or birth defects or other reproductive harm.

Please refer to [www.innovative-health.com](http://www.innovative-health.com) for product warranty.